Specification of Bi-PAP Machine

	·	Specification of Bi-PAP Machine
SN	Parameter	Specification
1	Usage	To aid in breathing during respiratory failure in COPD patient
2	Operational	Operable on AC and DC power supply and intended to be used as a bedside
	Requirements	device in ICU/HCU/Chest Ward.
3	System	Auto Bi-PAP Machine. Easily fitting silicone gel different sized face masks
2	Configuration	and head straps.
		IPAP 4 - 30 cm.
		EPAP 4 - 25 cm.
	Technical Specifications	Bi level pressure range: 3 - 25 cm H2O (0.5 cm H2O increments).
		Rise Time 100 - 600 msec.
4		Automatic Altitude Adjustment.
+		DC Operable
		ABOB analog output for PSG
		Automatic Leak Compensation
		Therapy Compliance Meter
		Auto on and off feature
		All standard accessories/ consumables/ parts should be included
		All standard maintenance tools and cleaning/lubrication materials should be
		included
		Travel bag (carrying cases)
5	Accessories, spares	Large slip - resistant pads on bottom of each unit to minimize movement on
	and consumables	nightstand.
		Medium sized easy fitting silk gel face mask
		Head strap.
		Humidifier chamber with over fill protection, corrosive resistant, SS plate,
		around 500ml capacity for 8hr use at the highest setting
6	Standard	ISO, CE or USFDA certified product certificate at the time of bidding.
		220 VAC, 50 Hz, fitted with appropriate power cable of at least 3 meter in
7	Power Supply	length.
		Inbuilt battery backup of more than 1 hour.
		The training should include hands on training to the user physician in the
	User Training	institution where the equipment is being used. It should also include the use
8		of all operational functions of the equipment, as well as routine checks and
		maintenance expected by users.
		Comprehensive warranty for 2 yrs and extra one year of AMC from the date
		of installation.
	Terms and	Authorized sales agent, trained engineer and after sales service for quoted
		model of machine must be locally and nationally available.
9		The equipment must be brand new
	Conditions	
		Date of manufacture should not be more than 2 yrs at the time of delivery.
		The supplier must accomplish proper installations and commissioning of the
		equipment on site.Free Installation.
	Maintenance/Servi ce during warranty period	During the warranty period the supplier must as
10		During the warranty period, the supplier must ensure preventive maintenance
		and corrective/ breakdown maintenance whenever required free of cost.
		Must provide user (Operating) manual in English.
11	Documentation	Must provide Service (Technical/Maintenance) manual in English.
		List of important spare parts and accessories with their part, number and
		cost Certificate of caliberation and inspection for the factory

cost. Certificate of caliberation and inspection for the factory.

Recommended ResMed, Philips, Devilbiss or equivalent. Company

		Specification of BMI Machine
S.N.	Parameters	Technical Specification
1	Usage	For measurement of Body Mass Index
	Features	Ultrasonic weight Measurement
		Precise Sensor Weight Measurement with BMI
		Fat Measurement & Body Composition
2		Blood Pressure & Pulse Measurement
		Infrared Body Temperature Measurement
		Operation Process Voice Prompt
		Multi -Media min10" digital Screen
3	Height measurement range	100-200 cm
4	Minimum division (height)	0.5 cm
5	Weight measurement range	8-200 kg
6	Minimum division (weight)	0.1 kg
7	Pulse	40-180 times/min
8	Power Supply	220 VAC, 50 Hz
9	Standard	ISO and CE Certified or equivalent
10	Warranty	Two years from date of installation
	Terms and conditions	Brand new
		Date of manufacture should be less than two years at the time of delivery
11		Should provide user manual in English
		Should provide service manual in English
		Free installation
		Installation is the responsibility of supplier
12	Recommended Company	Evana- X80, Sangong, DHM or equivalent

		Specifications of Stainless Steel Clip Boards
SN	Parameters	Technical Specifications
1	Usage	Stainless steel clipboards for easy document storage in hygiene conscious environments
	Feature	Material: 304 grade stainless steel construction
2		Can be easily clipped for storage of patient records
_		Complete with stainless steel clip to secure documents
		Size: A4 with wall hanging clip
3	Standards	CE/ ISO or equivalent
4	Terms and condition	Date of manufacture must be less than 2 years at the time of delivery, Brand New
5	Warranty	Minimum 1 year

Specifications of Finger tip Pulse Oximeter

	S.N	Parameters	Technical Specifications
+	1	Usage	To measure and display the oxygen saturation
~	2	Features	SPO2 & heart rated display, battery operated (rechargeable with charger and 1 set spare battery)

3	*3	Recommended Company or Brand	Dr.Morpen, devillbiss or equivalent
	4	Standards	CE / ISO or Equivalent
	5	Terms and Conditio	The device must be brand new and date of manufacture must be less than 2 years on day of delivery. Should provide rechargeable battery with charger set.
	6	Warranty	Minimum 1 year from the date of delivery

Specification of Stadiometer

S.N	Parameters	Specification
1	Usage	Measuring height of the patient
2	Item weight	Minimum 2.4 kg
	Features	Simple and easy to set up - no wall fastening neck
3		Large and sturdy floor plate to ensure stability
5		Result clearly visible while measuring
		Measuring range: 20 - 205 cm
4	Dimension	13.3"W x 85.2"H x 23.2"D (approximately)
5	Standard	ISO or equivalent Certified
6	Warranty	Minimum 2 years from date of delivery
7	Recommended	Seca 213 portable stadiometer (physical therapy aids), Indosurgicals
/	Company	height measuring scale, Prime surgicals or equivalent

Specification of X-ray View Box

S.N	Parameters	Technical Specification
1	Usages	To view the X-ray films
	Features	LED illumination system
		Acrylic Sheet
2		Fuse
4		Switch
		Dimmer
		Power Cord Socket
3	Dimension	One 14" x 17"
3		Width: 40-45cm
4	Shape	Rectangle
5	Color	White
6	System Power	200-240V, AC, 50Hz
	Terms and Conditions	ISO or CE approval
		Brand new
7		Date of manufacture should be not more than 2 years at time of delivery.
		Free Installation
8	Warranty	Minimum 2 years from the date of installation
	4	

Bushing the state him is the state of the st